....

Riverside Business Park – Block J Boulevard International 55 1070 Brussels, Belgium T. + 32 (0)2 528 74 00 F. + 32 (0)2 528 74 01

Email: info.europe@ansell.com

www.ansell.com



EU DECLARATION OF CONFORMITY

Manufacturer Name/Address:

Ansell Healthcare Europe NV

Boulevard International 55

1070 Brussels

Belgium

SRN Number:

BE-MF-000000691

Risk Class:

Class I

Intended Purpose:

A non-sterile medical device intended as an examination/treatment glove and a protective barrier when worn on the hands of healthcare providers, during patient examination, or for other sanitary purposes. The device is used mainly as a two-way barrier to protect both the patient and wearer against contaminants. This

is a single-use device.

EMDN Code and Description:

T01020204 - Nitrile Examination/Treatment Glove

Basic UDI-DI:

5414566 MF92134 DK

Product Name(s):

Product Name	Size	Product Code	Market Regions
Microflex® 92-134	XS	92134060	EMEA/INDIA
Microflex® 92-134	S	92134070	EMEA/INDIA
Microflex® 92-134	M	92134080	EMEA/INDIA
Microflex® 92-134	L	92134090	EMEA/INDIA
Microflex® 92-134	XL	92134100	EMEA/INDIA

Conformity Assessment Procedure:

Annex I & Annex II + Annex III

This EU Declaration of Conformity is issued under the sole responsibility of Ansell Healthcare Europe NV/SA.

We hereby declare that the medical device(s) specified above meet the provision of the Regulation (EU) MDR 2017/745 for medical devices.

_

Riverside Business Park – Block J Boulevard International 55 1070 Brussels, Belgium T. + 32 (0)2 528 74 00 F. + 32 (0)2 528 74 01

Email: info.europe@ansell.com www.ansell.com



Signed on behalf of Ansell Healthcare Europe NV

Name: Samantha Marshall

Ansell Healthcare Europe NV
Riverside Business Park - Block J
Bld Internationalelaan 55

B-1070 Brussels BELGIUM

Position: Samantila Marshall Position:

Director Regulatory Affairs Medical EMEA / APAC

Date of issue: 21 February 2022
Place of issue: Brussels, Belgium
Version No: MED\MFX92134\005