

EU DECLARATION OF CONFORMITY

Manufacturer Name/Address: Ansell Healthcare Europe NV/SA
Riverside Business Park,
Block J, Boulevard International 55,
1070 Brussels,
Belgium

SRN Number: BE-MF-000000691

Risk Class: Class I

Intended Purpose: A non-sterile medical device intended as an examination/treatment glove and a protective barrier when worn on the hands of healthcare providers, during patient examination, or for other sanitary purposes. The device is used mainly as a two-way barrier to protect both the patient and wearer against contaminants. This is a single-use device.

EMDN Code and Description: T01020204 - Nitrile Examination/Treatment Glove

Basic UDI-DI: 5414566 MF93833XCINT U3

Product Name(s):

Product Name	Size	Product Code	Market Regions
Microflex® 93-833	XS	93833060	EMEA/APAC
Microflex® 93-833	S	93833070	EMEA/APAC
Microflex® 93-833	M	93833080	EMEA/APAC
Microflex® 93-833	L	93833090	EMEA/APAC
Microflex® 93-833	XL	93833100	EMEA/APAC
Microflex® XCEED® XC-INT	XS	XC-INT-XS	EMEA/APAC
Microflex® XCEED® XC-INT	S	XC-INT-S	EMEA/APAC
Microflex® XCEED® XC-INT	M	XC-INT-M	EMEA/APAC
Microflex® XCEED® XC-INT	L	XC-INT-L	EMEA/APAC
Microflex® XCEED® XC-INT	XL	XC-INT-XL	EMEA/APAC

Conformity Assessment Procedure: Annex I & Annex II + Annex III

This EU Declaration of Conformity is issued under the sole responsibility of Ansell Healthcare Europe NV/SA.

We hereby declare that the medical device(s) specified above meet the provision of the Regulation (EU) MDR 2017/745 for medical devices.

Signed on behalf of Ansell Healthcare Europe NV



Ansell Healthcare Europe NV
Riverside Business Park - Block J
Bid Internationaalelaan 55
B-1070 Brussels
BELGIUM

Name: Samantha Marshall
Position: Director Regulatory Affairs Medical EMEA / APAC
Date of issue: 02 March 2023
Place of issue: Nuneaton, England
Version No: MED\MFX93833XC\002